



TRANSITION – Telescoping Knee Brace

Single Patient Use

Rx Only

IMPORTANT INFORMATION

Please read all instructions, warnings, and precautions before use. Correct application is essential for proper functioning of product.

INTENDED USE

This knee brace provides knee immobilization or range of motion limitations as prescribed by a medical practitioner. This brace may be used post-injury, post-operatively or throughout rehabilitation.

INDICATIONS

Post-operative rehabilitation

CONTRAINDICATIONS

Patients with poor peripheral circulation, diabetes or a decrease in skin sensitivity should not use this product.


PRECAUTIONS

This product is to be fitted by a physician (or properly licensed practitioner) who is familiar with the purpose for which they are responsible. The physician or practitioner is responsible for providing wearing instructions and precautions to other healthcare practitioners or healthcare providers involved in the patient's care and to the patient. Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.

INSTRUCTIONS FOR USE


Brace Application - Full Foam

1. Unhook all clip buckles and remove thigh and calf wraps.
2. Apply thigh and calf wraps with openings toward the




front. Excess wrap material can be trimmed.

3. Align the hinge of the brace with the knee joint. The brace arms will adhere to the foam wraps. Brace should fit 2" below groin and 2" above ankle bones. (See Length Adjustment.)
4. Fasten clip buckle just below the knee joint. Repeat with strap above the knee joint.
5. Next, fasten the two remaining straps beginning below the knee then above the knee.
6. Attach Y-tab closures on strap ends and tighten all four straps. Straps should be snug but comfortable. Excess strapping can be trimmed before applying Y-tabs.



Brace Application – Cool Foam

1. Unhook all clip buckles.
2. Align the hinge of the brace with the knee joint. Brace should fit 2" below groin and 2" above ankle bones. (See Length Adjustment.)
3. Fasten clip buckle just below the knee joint. Repeat with strap above the knee joint.
4. Next, fasten the two remaining straps beginning below the knee then above the knee.
5. Attach Y-tab closures on strap ends and tighten all four straps. Straps should be snug but comfortable. Excess strapping can be trimmed before applying Y-tabs.



TRANSITION – Telescoping Knee Brace

Single Patient Use

Rx Only

IMPORTANT INFORMATION

Please read all instructions, warnings, and precautions before use. Correct application is essential for proper functioning of product.

INTENDED USE

This knee brace provides knee immobilization or range of motion limitations as prescribed by a medical practitioner. This brace may be used post-injury, post-operatively or throughout rehabilitation.

INDICATIONS

Post-operative rehabilitation

CONTRAINDICATIONS

Patients with poor peripheral circulation, diabetes or a decrease in skin sensitivity should not use this product.


PRECAUTIONS

This product is to be fitted by a physician (or properly licensed practitioner) who is familiar with the purpose for which they are responsible. The physician or practitioner is responsible for providing wearing instructions and precautions to other healthcare practitioners or healthcare providers involved in the patient's care and to the patient. Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.

INSTRUCTIONS FOR USE

Brace Application - Full Foam

1. Unhook all clip buckles and remove thigh and calf wraps.
2. Apply thigh and calf wraps with openings toward the



front. Excess wrap material can be trimmed.

3. Align the hinge of the brace with the knee joint. The brace arms will adhere to the foam wraps. Brace should fit 2" below groin and 2" above ankle bones. (See Length Adjustment.)
4. Fasten clip buckle just below the knee joint. Repeat with strap above the knee joint.
5. Next, fasten the two remaining straps beginning below the knee then above the knee.
6. Attach Y-tab closures on strap ends and tighten all four straps. Straps should be snug but comfortable. Excess strapping can be trimmed before applying Y-tabs.



Brace Application – Cool Foam

1. Unhook all clip buckles.
2. Align the hinge of the brace with the knee joint. Brace should fit 2" below groin and 2" above ankle bones. (See Length Adjustment.)
3. Fasten clip buckle just below the knee joint. Repeat with strap above the knee joint.
4. Next, fasten the two remaining straps beginning below the knee then above the knee.
5. Attach Y-tab closures on strap ends and tighten all four straps. Straps should be snug but comfortable. Excess strapping can be trimmed before applying Y-tabs.






Length Adjustment

1. Unlock length adjustment buttons located at the top and bottom on both sides of brace.
2. Adjust to accommodate leg length by sliding bars. Brace should fit 2" below groin and 2" above ankle bones.
3. Lock length adjustment buttons ensuring each side is equal in length.




Flexion/Extension Adjustment

1. The clear plastic dial can be rotated to expose flexion or extension pin.
2. Pull pin outward and slide to desired location.
3. Reseat pin back into place.
4. Perform function on both hinges.



Drop-Lock Adjustment

1. To completely lock leg into full extension, push blue lever into the down position.
2. To unlock, pull blue lever into the up and locked position.



WASHING INSTRUCTIONS

Hand wash liner with mild detergent in lukewarm water. Air dry.



Distributed by DeRoyal
200 DeBusk Lane
Powell, TN 37849 U.S.A.
888.938.7828
865.938.7828

Part #0-KB7000
©2010 DeRoyal. All Rights Reserved.
Revised 1/10

MediRoyal Nordic AB
Staffans väg 6 B,
S-192 07 Sollentuna
SWEDEN
+46 8 506 766 00
www.mediroyal.se




Length Adjustment

1. Unlock length adjustment buttons located at the top and bottom on both sides of brace.
2. Adjust to accommodate leg length by sliding bars. Brace should fit 2" below groin and 2" above ankle bones.
3. Lock length adjustment buttons ensuring each side is equal in length.




Flexion/Extension Adjustment

1. The clear plastic dial can be rotated to expose flexion or extension pin.
2. Pull pin outward and slide to desired location.
3. Reseat pin back into place.
4. Perform function on both hinges.



Drop-Lock Adjustment

1. To completely lock leg into full extension, push blue lever into the down position.
2. To unlock, pull blue lever into the up and locked position.



WASHING INSTRUCTIONS

Hand wash liner with mild detergent in lukewarm water. Air dry.



Distributed by DeRoyal
200 DeBusk Lane
Powell, TN 37849 U.S.A.
888.938.7828
865.938.7828

Part #0-KB7000
©2010 DeRoyal. All Rights Reserved.
Revised 1/10

MediRoyal Nordic AB
Staffans väg 6 B,
S-192 07 Sollentuna
SWEDEN
+46 8 506 766 00
www.mediroyal.se